

校准证书

CALIBRATION CERTIFICATE

证书编号:



第 1 页 共 3 页

Certificate No.

J202107020134A-0007-G1

Page of

委托方

中南大学湘雅三医院

Client

联络信息

长沙市岳麓区桐梓坡路

Contact Inf.

仪器名称

无线温湿度变送器

Description

型号/规格

RCW-360

制造厂

精创

Model/Type

Manufacturer

出厂编号

13006 24801 51813 75433

管理号

614170

Serial No.

Asset No.

接收日期

2021年07月14日

校准日期

2021年07月14日

Receipt Date

Y M D

Cal. Date

Y M D

发布日期

2021年07月14日

Issued Date

Y M D

批准

Approved by

周焰华 (副主任)

审核

Inspected by

莫丽娟

校准

Calibrated by

黄蛟

证书专用章

(Stamp)

总部地址(Headquarters Add):广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd, West of HuangPu Ave.Guangzhou.Guangdong .China

实验室地址(Add.of the Lab): 广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd,West of HuangPu Ave.Guangzhou,Guangdong,China

联系电话(Tel.):400-602-0999

邮政编码(Postcode):510656

网站(Website):http:// www.grgtest.com

电子邮件(E-mail):grgtest@grgtest.com



扫一扫验真伪

校 准 说 明
DIRECTIONS OF CALIBRATION

证书编号: J202107020134A-0007-G1

第 2 页 共 3 页

Certificate No.

Page of

- 1.本实验室的质量管理体系符合ISO/IEC 17025:2017标准的要求。
(The quality system is in accordance with ISO/IEC 17025:2017.)
- 2.本结果仅对本次校准样品有效。未经实验室批准,不得部分复制。如有疑问请在15个工作日内反馈。
(The result is only valid for the calibrated sample.The certificate shall not be reproduced except in full,without the written approval of our laboratroy .please feedback to us within 15 days if you have any question.)
3. 本证书编号具有唯一性, 后缀若带有“-Gx”的证书为替换证书, 自发出后原证书即刻作废。
(Each certificate has a unique number. The suffix of "-Gx" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.)
- 4.证书中最大允许误差、判定结果仅供参考, 其中“P”代表“合格”, “F”代表“不合格”, “N/A”代表“不适用”。使用人员应结合实际测量需求, 评估测量不确定度对符合性评定的影响。(MPE & judgement result in the datasheet is only for reference , "P" is "Pass" , "F" is "Fail" and "N/A" is "Not Applicable".Whereas users should evaluate the effects of MU of calibration results on conformance assessment by actual measurement.)
- 5.本次校准的技术依据(Reference document for calibration)。
JJG(苏)99-2010数字温湿度计检定规程(V.R of Digital Thermo-hygrometers)

6. 本次校准使用的主要测量标准(Main Standards of Measurement Used in the Calibration.):

名称	编 号	证书号/有效期	溯源机构	技术特征
Description	Serial No.	Certificate No./ Due Date	Traceability Institute	Technique Character
无线温湿度记录仪	A15120871~A151 20880	J202102022489- 0009 2022-02-03	广州广电计量检 测股份有限公司	温度: $\pm 0.3^{\circ}\text{C}$, 湿度: $\pm 2.0\%$ RH

7.校准地点、环境条件(Place and environmental conditions of the calibration):

地点	温度	18.6	$^{\circ}\text{C}$	相对湿度	61	%
Place	客户外科楼1楼	Temperature		Relative Humidity		

8. 建议复校时间间隔: 1 年, 送校单位也可按实际使用情况自主决定。

Suggested calibration interval is 1 year or it can be altered depending on the actual usage of the user.

量检测



报告专
(01)

校准结果
RESULTS OF CALIBRATION

证书编号: J202107020134A-0007-G1

第 3 页 共 3 页

Certificate No.

Page of

一、外观: 符合要求

Appearance: Pass

二、相对湿度示值校准: (%RH)

Relative Humidity Indicating Value Calibration:

标准示值 Ref.Value	被检示值 Indicated	被检器修正值 Correction	不确定度 $U(k=2)$
53.9	49	+5	3

三、温度示值校准: (°C)

Temperature Indicating Value Calibration:

标准示值 Ref.Value	被检示值 Indicated	被检器修正值 Correction	不确定度 $U(k=2)$
18.6	17.2	+1.4	0.4

备注:

Notes:

结论(Conclusion): 按校准结果使用

1.本报告中的扩展不确定度是由标准不确定度乘以包含概率约为95%时的包含因子 k 。The expanded uncertainty is given in the report by the standard uncertainty multiplied by the probability of about 95% when the factor k .

2.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

(以下空白)

(The below is blank)

