

湖南省计量检测研究院

Hunan Institute of Metrology and Test

检定证书

Verification Certificate

证书编号:



Certificate No. 2024071618593005

委托单位 中南大学湘雅三医院

Applicant

计量器具名称 医用超声诊断仪

Name of Instrument

型号/规格 EPIQ 7

Type/Specification

出厂编号 US419B0594

Serial No.

制造单位 PHILIPS

Manufacturer

检定依据 JJG 639-1998

Verification Regulation

检定结论 A档

Conclusion

批准人
Approved by

张秦

张秦

核验员

Checked by

巢世友

巢世友

检定员

Verified by

刘融峰

刘融峰

检定日期

Date of Verification

有效期至

Valid Date

2024 年 07 月 03 日

Year Month Day

2025 年 07 月 02 日

Year Month Day

地址: 长沙市香樟路396号

Address

邮编: 410014

Post Code

电话/传真: 0731-85581751/85687602

Telephone

电子邮件: jszb@hnjly.cn

Email

证书编号： 2024071618593005



我院系国家法定计量检定机构

Our Institute is National Legal Institute of Verification

计量授权机构：国家市场监督管理总局

Authorization Body: State Administration for Market Regulation

计量授权证书号：（国）法计（2022）01029号

Authorized certificate No.

测量溯源性说明：本次检定使用的计量器具均可溯源到国家基准

Statement of measurement traceability :All measuring instruments used in the verification can be traced back to national standards

检定所使用的计量标准：

Main equipments of measurement used in the verification

名称 Name	测量范围 Type/Specification	不确定度/准确度等级 Uncertainty/Accuracy class	标准证书号 Serial No	有效期至 Valid Date
医用超声诊断仪超声源检定装置	超声功率2mW~20W	$U_{rel} = 10\% (k = 2)$	[2004]国量标湘证字第078号	2028-05-15

检定地点及其环境条件：

Address and environmental condition in the verification

地点： 中南大学湘雅三医院（超声科）

Address

温度： 26 °C 相对湿度： 60 % 其他： /
Temperature Relative humidity else

限制使用条件和测量范围： /

Limited conditions and measuring range

注 (Note):

- 我院仅对加盖“湖南省计量检测研究院检定专用章”的完整证书负责。
Our Institute is only responsible to the complete certificates stamped with "Hunan Institute of Metrology seal"
- 本证书的检定结果仅对所检定计量器具有效。
The certificates are only to be effective for the instruments examined by our Institute
- 请妥善保管此证书。
Please properly keep this certificate
- 本证书中英文两种语言表达，准确含义以中文为准。
The Certificate is reported in both English and Chinese, with the Chinese version as standard.





检 定 结 果
Verification results

1. 外观及工作正常性： 符合。
2. 探头（标称频率：3.0MHz，扫描方式：凸阵）

序号	检定项目	技术要求	检定结果			
1	探测深度	$\geq 160\text{mm}$	240mm			
2	盲区	$\leq 8\text{mm}$	4mm			
3	侧、轴向分辨力	/	靶群深度(mm)	30	70	120
		$\leq 2\text{mm}$ (深度 $\leq 80\text{mm}$) $\leq 4\text{mm}$ ($80\text{mm} < \text{深度} \leq 130\text{mm}$)	侧向(mm)	1	2	3
		$\leq 1\text{mm}$ (深度 $\leq 80\text{mm}$) $\leq 2\text{mm}$ ($80\text{mm} < \text{深度} \leq 130\text{mm}$)	轴向(mm)	1	1	2
4	几何位置示值误差	纵向 $\leq 10\%$ ；横向 $\leq 10\%$	纵向：1.0 %	横向：2.5 %		
5	囊性病灶直径误差	$\pm 10\%$	纵向：-1.0 %	横向：-3.0 %		



以下空白